



校准证书

CALIBRATION CERTIFICATE

证书编号 DYQ202001352
Certificate No.

第 1 页, 共 3 页
Page of

委托方 _____
Client

委托方联络信息 _____
Contact Information

计量器具名称 钳形接地电阻仪 (钳形接地电阻测试仪)
Description

型号/规格 ETCR2000+
Model/Type

制造厂 广州市铨泰电子科技有限公司
Manufacturer

出厂编号 20500369 设备管理编号 _____
Serial No. Equipment No.

接收日期 2020 年 05 月 26 日
Date of Receipt Y M D

结果 见校准结果
Results Shown in the results of calibration

校准日期 2020 年 06 月 03 日
Date of Calibration Y M D

批准人 张剑 张剑
Approved Signatory

核 验 李华杰 李华杰
Reviewed by

校 准 庞俊南 庞俊南
Calibrated by

证书专用章
Stamp



扫一扫查真伪

本中心地址: 中国广州市广园中路松柏东街30号 邮政编码: 510405
电话: (8620)86594172 传真: (8620)86590743 投诉电话: (8620)36611242 E-mail: scm@scm.com.cn

Add: No.30, Songbai East Street, Guangyuan Middle Road, Guangzhou, Guangdong, China
Post Code: 510405 Tel: (8620)86594172 Fax: (8620)86590743 Complaint Tel: (8620)36611242

证书真伪查询: www.scn.com.cn; www.mtppsp.com Certificate Authenticity Identify: www.scn.com.cn; www.mtppsp.com



说 明

证书编号 DYQ202001352
Certificate No.

DIRECTIONS

第 2 页, 共 3 页
Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the State Administration for Market Regulation. The quality system is in accordance with ISO/IEC 17025:2017.

2. 本中心所出具的数据均可溯源至国家计量基准和国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and International System of Units (SI).

3. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 本院电磁实验室	温度 (20±3) °C	相对湿度 (40~55) %
Place (Electrics-magnetics Lab)	Temperature	R.H.

4. 本次校准的技术依据:

Reference documents for the calibration:

JJG1054-2009 钳形接地电阻仪检定规程 V. R. of Clamp Earth Resistance Meters

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格 Name of Equipment /Model/Type	编号 Serial No.	证书号/有效期/溯源单位 Certificate No./Due Date /Traceability to	计量特性 Metrological Characteristic
实验室直流电阻器 DC Resistance Box /ZX54	9208293	DYQ201906943 /2020-12-15 /本中心	0.01 级 Grade 0.01

注: 1. 本证书校准结果只与受校准仪器有关。The results relate only to the items calibrated.

Note: 2. 未经本机构书面批准, 不得部分复制此证书。This certificate shall not be reproduced except in full, without the written approval of our laboratory.

3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注。The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items.

4. 本次校准日期视为发布日期。The calibration date is the date of issue of the certificate.



校准结果

RESULTS OF CALIBRATION

证书编号 DYQ202001352
Certificate No.

原始记录号 020201352
Record No.

第 3 页, 共 3 页
Page of

1 外观检查 (Apparent inspection): 符合要求 (Pass)

2 电阻测量见表1(Resistance test shown in table 1):

表1(Table 1)

示值 Indication	标准值 Reference Value	误差 Error	允许误差 MPE	结论 Conclusion
(Ω)	(Ω)	(Ω)	(Ω)	(P/F)
0.10	0.100	0.00	± 0.01	P
0.20	0.200	0.00	± 0.01	P
0.50	0.500	0.00	± 0.02	P
1.0	1.00	0.0	± 0.1	P
2.0	2.00	0.0	± 0.1	P
5.0	5.00	0.0	± 0.2	P
10.0	10.0	0.0	± 0.2	P
20.0	20.0	0.0	± 0.3	P
49.9	50.0	- 0.1	± 1.3	P
99.9	100	- 0.1	± 3.0	P
198	200	- 2	± 15	P
490	500	- 10	± 60	P
780	800	- 20	± 180	P
960	1000	- 40	± 280	P

说明:

Note:

1 本次测量结果的扩展不确定度:

The Expanded Uncertainty of Measurement:

电阻 $U_{rel}=0.6\%$ 包含因子 $k=2$

Resistance Coverage factor

本证书中给出的扩展不确定度依据JJF1059.1-2012《测量不确定度评定与表示》评定,由合成标准不确定度乘以包含概率约为95%时对应的包含因子 k 得到。

The expanded uncertainty given in this certificate is evaluated according to JJF1059.1-2012 "Evaluation and Expression of Uncertainty in Measurement", which is obtained by multiplying the combined standard uncertainty by the coverage factor k corresponding to the coverage probability of about 95%.

2 按照所依据技术文件的规定, 建议复校时间间隔不超过1年。

According to the demand of reference document, next calibration is proposed within 1 year.